



Report Number:	23-003614/D003.R000
Report Date:	04/10/2023
ORELAP#:	OR100028
Purchase Order:	
Received:	04/04/23 13:38

Customer:	IHC LLC
Product identity:	01FLT_SourChem
Client/Metrc ID:	
Laboratory ID:	23-003614-0002

Summ	nary
------	------

Analyte	Result (%)			
CBD-A	19.8		CBD-Total	18.7%
CBD	1.36	• CBD-A		
CBG-A	0.570	CBD	THC-Total	<loq< td=""></loq<>
CBDV-A	0.129	CBG-A		
CBG	0.117	CBDV-A	(Reported in pe	rcent of total sample)
CBC-A	0.0338	CBG		
		CBC-A		

Page 1 of 9
Test results relate only to the parameters tested and to the samples as received by the laboratory. Test results meet all requirements of NELAP and the Columbia Laboratories quality assurance plan
unless otherwise noted. This report shall not be reproduced, except in full, without the written consent of this laboratory. Samples will be retained for a maximum of 30 days from the receipt date unless
prior arrangements have been made.
Testing in accordance with: OAR 333-007-0430



IHC LLC

.

No 17.4 °C

Ramos

825 NW 16th Ave Portland Oregon 97209

01FLT_SourChem

23-003614-0002

United States of America (USA)

Customer:

Product identity:

Client/Metrc ID:

Sample Date:

Laboratory ID:

Temp:

Evidence of Cooling:

Relinquished by:

12423 NE Whitaker Way Portland, OR 97230 503-254-1794



Report Number:	23-003614/D003.R000
Report Date:	04/10/2023
ORELAP#:	OR100028
Purchase Order:	
Received:	04/04/23 13:38



Sample Results

Potency		2015 V98-6 (mod) [‡]	Units %	Batch: 2305753	Analyze: 4/5/23 7:29:00 PM
Analyte			Notes		
000	Received	-			
CBC	< LOQ	0.0292			• CBD-A
CBC-A	0.0338	0.0292			• CBD-A • CBD
CBC-Total	< LOQ	0.0548			• CBD • CBG-A
CBD	1.36	0.0292			O CBDV-A
CBD-A	19.8	0.292			• CBG
CBD-Total	18.7	0.285			O CBC-A
CBDV	< LOQ	0.0292			
CBDV-A	0.129	0.0292			
CBDV-Total	0.112	0.0545			
CBE	< LOQ	0.0292			
CBG	0.117	0.0292			
CBG-A	0.570	0.0292			
CBG-Total	0.617	0.0545			
CBL	< LOQ	0.0292			
CBL-A	< LOQ	0.0292			
CBL-Total	< LOQ	0.0548			
CBN	< LOQ	0.0292			
CBT	< LOQ	0.0292			
∆10-THC-9R	< LOQ	0.0292			
∆10-THC-9S	< LOQ	0.0292			
$\Delta 10$ -THC-Total	< LOQ	0.0584			
∆8-THC	< LOQ	0.0292			
∆8-THCV	< LOQ	0.0292			
∆9-THC	< LOQ	0.0292			
exo-THC	< LOQ	0.0292			
THC-A	< LOQ	0.0292			
THC-Total	< LOQ	0.0548			
THCV	< LOQ	0.0292			
THCV-A	< LOQ	0.0292			
THCV-Total	< LOQ	0.0545			
Total Cannabinoids	22.0				

www.columbialaboratories.com

Page 2 of 9

Test results relate only to the parameters tested and to the samples as received by the laboratory. Test results meet all requirements of NELAP and the Columbia Laboratories quality assurance plan unless otherwise noted. This report shall not be reproduced, except in full, without the written consent of this laboratory. Samples will be retained for a maximum of 30 days from the receipt date unless prior arrangements have been made.



12423 NE Whitaker Way Portland, OR 97230 503-254-1794



23-003614/D003.R000 **Report Number: Report Date:** 04/10/2023 ORELAP#: OR100028 Purchase Order: Received: 04/04/23 13:38

 www.columbialaboratories.com
 Page 3 of 9

 Test results relate only to the parameters tested and to the samples as received by the laboratory. Test results meet all requirements of NELAP and the Columbia Laboratories quality assurance plan unless otherwise noted. This report shall not be reproduced, except in full, without the written consent of this laboratory. Samples will be retained for a maximum of 30 days from the receipt date unless prior arrangements have been made.

 Testing in accordance with:
 OAR 333-007-0430





 Report Number:
 23-003614/D003.R000

 Report Date:
 04/10/2023

 ORELAP#:
 OR100028

 Purchase Order:
 Received:

 04/04/23 13:38
 04/04/23 13:38

Abbreviations

Limits: Action Levels per OAR-333-007-0400, OAR-333-007-0210, OAR-333-007-0220, CCR title 16-division 42. BCC-section 5723

Limit(s) of Quantitation (LOQ): The minimum levels, concentrations, or quantities of a target variable (e.g., target analyte) that can be reported with a specified degree of confidence.

^b = ISO/IEC 17025:2017 accredited method.

Units of Measure

% = Percentage of sample % wt = $\mu g/g$ divided by 10,000

Approved Signatory

Derrick Tanner General Manager

www.columbialaboratories.com

Page 4 of 9

Test results relate only to the parameters tested and to the samples as received by the laboratory. Test results meet all requirements of NELAP and the Columbia Laboratories quality assurance plan unless otherwise noted. This report shall not be reproduced, except in full, without the written consent of this laboratory. Samples will be retained for a maximum of 30 days from the receipt date unless prior arrangements have been made.

Columbia
LABORATORIES
💽 A Tentamus Company

騣	
	2

Report Number:	23-003614/D003.R000
Report Date:	04/10/2023
ORELAP#:	OR100028

Purchase Order:

Received:

04/04/23 13:38

					117.4	. 1	italya	in Rea	ueste	ud .					0 Burnhar	
Ihe Hemp Collect Some: All NW Flanders s Portland Same: Grail Results dropbox (IH (61) 608164 PAResult get @thehei	UP np (97209	editidae - OR 35 compounds	stide Matthesdar - 379 compositin		chad Solversa	oliture & Water Activity		cro: Yeart and MoM	ore-2.06 and Total Coliform	49			Projec Proj Castorn P Report to	nt Number: Int Name: Importing: Inf time: Inf time:	ETRE or C Other Excloses Day Randard Turnaround Excloses Day Rish Turnaround* Excloses Day Rish Turnaround* Excloses Day Rish Turnaround* "C'ext for evolution
Clean Senate Menification D1FLT LamonMeringu	Oute	Time .	Pedicida	Pessidie	X Folieres	factual 3	Molettore	Tarpenes	Micros: Ve	Micro-4.0	Harry Mutals	Mycotcali	Offer:	Sarapie Type 1	Weight (Unita)	Common ta/Metric 10
01FLT_Canonimeningu	-		_	-	X	-			_	-	-	_		v		
01FLT_SourHH	-		_	-	X	-	_	-	-	-	-	_		v		
09ATLIRDST_In_CML					×									ċ		
			_	-	-							-				
					-											
		Time -			-	(inver			-		if wi	-	110			Lab Use Only.

Y - Sateple Type Codes: Vegetation (V); holates (I); Edites (I); Edites (I); Hecture/Tepical (I); Edites (I); Beverage (I);

Respice table(1) by Calmabia Laborateine with justing regimented consistent an agree NAC NO. tas a accession with the consetures of service associated and its COC. In signing "Advanced of the "prevair agreeing webcan innor

13423 Mt Whiteker Way Partiand, OR 97230

A: (800) 254-2784 (Amr. (800) 254-3462 Info@unicellandeousles.com

Non of

Page 5 of 9 <u>www.columbialaboratories.com</u> Test results relate only to the parameters tested and to the samples as received by the laboratory. Test results meet all requirements of NELAP and the Columbia Laboratories quality assurance plan unless otherwise noted. This report shall not be reproduced, except in full, without the written consent of this laboratory. Samples will be retained for a maximum of 30 days from the receipt date unless prior arrangements have been made. Test results meet all requirements of NELAP and the Columbia Laboratories quality assurance plan prior arrangements have been made.





Report Number:	23-003614/D003.R000
Report Date:	04/10/2023
ORELAP#:	OR100028
Purchase Order:	
Received:	04/04/23 13:38

Revision: 1 Document ID: 7148 Legacy ID: Worksheet Validated 04/20/2021

			Lal	boratory	/ Quality Co					
J AOAC 2015 V98-6					В	atch ID:	23057	753		
Laboratory Control										
Analyte	LCS	Result	Spike	Units	% Rec		Limits		Evaluation	Notes
CBDVA	2	0.0336	0.033	%	101	80.0	-	120	Acceptable	
CBDV	2	0.0326	0.033	%	97.8	80.0	-	120	Acceptable	
CBE	2	0.0335	0.033	%	100	80.0	-	120	Acceptable	
CBDA	1	0.0300	0.031	%	97.0	90.0	-	110	Acceptable	
CBGA	1	0.0249	0.026	%	96.8	80.0	-	120	Acceptable	
CBG	1	0.0306	0.031	%	98.2	80.0	-	120	Acceptable	
CBD	1	0.0265	0.027	%	96.6	90.0	-	110	Acceptable	
THCV	2	0.0342	0.033	%	102	80.0	-	120	Acceptable	
d8THCV	2	0.0327	0.033	%	98.1	80.0	-	120	Acceptable	
THCVA	2	0.0336	0.033	%	101	80.0	-	120	Acceptable	
CBN	1	0.0269	0.027	%	99.5	80.0	-	120	Acceptable	
exo-THC	2	0.0318	0.033	%	95.5	80.0	-	120	Acceptable	
d9THC	1	0.0325	0.031	%	104	90.0	-	110	Acceptable	
d8THC	1	0.0312	0.031	%	100.0	90.0	-	110	Acceptable	
9S-d10THC	1	0.0318	0.031	%	101	80.0	-	120	Acceptable	
CBL	2	0.0344	0.033	%	103	80.0	-	120	Acceptable	
9R-d10THC	1	0.0297	0.032	%	93.0	80.0	-	120	Acceptable	
CBC	2	0.0332	0.033	%	99.6	80.0	-	120	Acceptable	
THCA	1	0.0350	0.036	%	97.4	90.0	-	110	Acceptable	
CBCA	2	0.0340	0.033	%	102	80.0	-	120	Acceptable	
CBLA	2	0.0343	0.033	%	103	80.0	-	120	Acceptable	
CBT	2	0.0320	0.033	%	96.0	80.0	-	120	Acceptable	
Method Blank	_									
Analyte CBDVA		esult	LOQ 0.03	-	Units %		Limits	•	Evaluation Acceptable	Notes
CBDVA		LOQ LOQ	0.03		%		0.03		Acceptable	
CBE		LOQ LOQ	0.03		%		0.03		Acceptable	
CBDA		LOQ LOQ	0.03		%		0.03		Acceptable	
CBGA		LOQ LOQ	0.03		%				Acceptable	
CBG		LOQ LOQ	0.03		%		< 0.03 < 0.03		Acceptable	
CBD		LOQ	0.03	-	%		0.03		Acceptable	
THCV		100	0.03		%				Acceptable	
d8THCV		LOQ LOQ	0.03		%		< 0.03		Acceptable	
THCVA		LOQ LOQ	0.03		%		0.03		Acceptable	
CBN		100	0.03		%		0.03		Acceptable	
exo-THC		LOQ LOQ	0.03		%		0.03		Acceptable	
dothc		LOQ LOQ	0.03		%		0.03		Acceptable	
d8THC		LOQ LOQ	0.03		%		0.03		Acceptable	
9S-d10THC		100	0.03		%		0.03		Acceptable	
CBL		LOQ LOQ	0.03		%		0.03		Acceptable	
9R-d10THC		LOQ	0.03		%		0.03		Acceptable	
CBC		LOQ LOQ	0.03		%		0.03		Acceptable	
THCA			0.03		%		0.03		Acceptable	
CBCA		LOQ LOQ	0.03		%		0.03		Acceptable	
CBLA		LOQ LOQ	0.03		%		0.03		Acceptable	
CBLA		LOQ LOQ	0.03		%		0.03		Acceptable	
Abbraviations	<pre></pre>	LUQ	0.05	1	/0	1 1	0.05		мссерсаріе	

Abbreviations

ND - None Detected at or above MRL RPD - Relative Percent Difference

LOQ - Limit of Quantitation

Units of Measure: % - Percent

Page 6 of 9 Test results relate only to the parameters tested and to the samples as received by the laboratory. Test results meet all requirements of NELAP and the Columbia Laboratories quality assurance plan
unless otherwise noted. This report shall not be reproduced, except in full, without the written consent of this laboratory. Samples will be retained for a maximum of 30 days from the receipt date unless
prior arrangements have been made.
Testing in accordance with: OAR 333-007-0430





Report Number:	23-003614/D003.R000	
Report Date:	04/10/2023	
ORELAP#:	OR100028	
Purchase Order:		
Received:	04/04/23 13:38	

Revision: 1 Document ID: 7148 Legacy ID: Worksheet Validated 04/20/2021

J AOAC 2015 V98-6					Bat	ch ID: 2305753			
Sample Duplicate	Sample ID: 23-003046-0001								
Analyte	Result	Org. Result	LOQ	Units	RPD	Limits	Evaluation	Notes	
CBDVA	0.0688	0.0697	0.03	%	1.25	< 20	Acceptable		
CBDV	<loq< td=""><td><loq< td=""><td>0.03</td><td>%</td><td>NA</td><td>< 20</td><td>Acceptable</td><td></td></loq<></td></loq<>	<loq< td=""><td>0.03</td><td>%</td><td>NA</td><td>< 20</td><td>Acceptable</td><td></td></loq<>	0.03	%	NA	< 20	Acceptable		
CBE	<loq< td=""><td><loq< td=""><td>0.03</td><td>%</td><td>NA</td><td>< 20</td><td>Acceptable</td><td></td></loq<></td></loq<>	<loq< td=""><td>0.03</td><td>%</td><td>NA</td><td>< 20</td><td>Acceptable</td><td></td></loq<>	0.03	%	NA	< 20	Acceptable		
CBDA	12.7	12.8	0.03	%	0.503	< 20	Acceptable		
CBGA	0.156	0.185	0.03	%	17.0	< 20	Acceptable		
CBG	<loq< td=""><td><loq< td=""><td>0.03</td><td>%</td><td>NA</td><td>< 20</td><td>Acceptable</td><td></td></loq<></td></loq<>	<loq< td=""><td>0.03</td><td>%</td><td>NA</td><td>< 20</td><td>Acceptable</td><td></td></loq<>	0.03	%	NA	< 20	Acceptable		
CBD	0.667	0.675	0.03	%	1.15	< 20	Acceptable		
THCV	<loq< td=""><td><loq< td=""><td>0.03</td><td>%</td><td>NA</td><td>< 20</td><td>Acceptable</td><td></td></loq<></td></loq<>	<loq< td=""><td>0.03</td><td>%</td><td>NA</td><td>< 20</td><td>Acceptable</td><td></td></loq<>	0.03	%	NA	< 20	Acceptable		
d8THCV	0.110	0.112	0.03	%	2.08	< 20	Acceptable		
THCVA	<loq< td=""><td><loq< td=""><td>0.03</td><td>%</td><td>NA</td><td>< 20</td><td>Acceptable</td><td></td></loq<></td></loq<>	<loq< td=""><td>0.03</td><td>%</td><td>NA</td><td>< 20</td><td>Acceptable</td><td></td></loq<>	0.03	%	NA	< 20	Acceptable		
CBN	<loq< td=""><td><loq< td=""><td>0.03</td><td>%</td><td>NA</td><td>< 20</td><td>Acceptable</td><td></td></loq<></td></loq<>	<loq< td=""><td>0.03</td><td>%</td><td>NA</td><td>< 20</td><td>Acceptable</td><td></td></loq<>	0.03	%	NA	< 20	Acceptable		
exo-THC	<loq< td=""><td><loq< td=""><td>0.03</td><td>%</td><td>NA</td><td>< 20</td><td>Acceptable</td><td></td></loq<></td></loq<>	<loq< td=""><td>0.03</td><td>%</td><td>NA</td><td>< 20</td><td>Acceptable</td><td></td></loq<>	0.03	%	NA	< 20	Acceptable		
d9THC	<loq< td=""><td><loq< td=""><td>0.03</td><td>%</td><td>NA</td><td>< 20</td><td>Acceptable</td><td></td></loq<></td></loq<>	<loq< td=""><td>0.03</td><td>%</td><td>NA</td><td>< 20</td><td>Acceptable</td><td></td></loq<>	0.03	%	NA	< 20	Acceptable		
d8THC	25.6	25.7	0.03	%	0.312	< 20	Acceptable		
9S-d10THC	<loq< td=""><td><loq< td=""><td>0.03</td><td>%</td><td>NA</td><td>< 20</td><td>Acceptable</td><td></td></loq<></td></loq<>	<loq< td=""><td>0.03</td><td>%</td><td>NA</td><td>< 20</td><td>Acceptable</td><td></td></loq<>	0.03	%	NA	< 20	Acceptable		
CBL	<loq< td=""><td><loq< td=""><td>0.03</td><td>%</td><td>NA</td><td>< 20</td><td>Acceptable</td><td></td></loq<></td></loq<>	<loq< td=""><td>0.03</td><td>%</td><td>NA</td><td>< 20</td><td>Acceptable</td><td></td></loq<>	0.03	%	NA	< 20	Acceptable		
9R-d10THC	<loq< td=""><td><loq< td=""><td>0.03</td><td>%</td><td>NA</td><td>< 20</td><td>Acceptable</td><td></td></loq<></td></loq<>	<loq< td=""><td>0.03</td><td>%</td><td>NA</td><td>< 20</td><td>Acceptable</td><td></td></loq<>	0.03	%	NA	< 20	Acceptable		
CBC	0.0627	<loq< td=""><td>0.03</td><td>%</td><td>NA</td><td>< 20</td><td>Acceptable</td><td></td></loq<>	0.03	%	NA	< 20	Acceptable		
THCA	0.505	0.511	0.03	%	1.05	< 20	Acceptable		
CBCA	0.584	0.600	0.03	%	2.60	< 20	Acceptable		
CBLA	<loq< td=""><td><loq< td=""><td>0.03</td><td>%</td><td>NA</td><td>< 20</td><td>Acceptable</td><td></td></loq<></td></loq<>	<loq< td=""><td>0.03</td><td>%</td><td>NA</td><td>< 20</td><td>Acceptable</td><td></td></loq<>	0.03	%	NA	< 20	Acceptable		
CBT	0.0820	0.0754	0.03	%	8.42	< 20	Acceptable		

ND - None Detected at or above MRL RPD - Relative Percent Difference LOQ - Limit of Quantitation

Units of Measure:

% - Percent

Page 7 of 9 Test results relate only to the parameters tested and to the samples as received by the laboratory. Test results meet all requirements of NELAP and the Columbia Laboratories quality assurance plan
unless otherwise noted. This report shall not be reproduced, except in full, without the written consent of this laboratory. Samples will be retained for a maximum of 30 days from the receipt date unless
prior arrangements have been made.
Testing in accordance with: OAR 333-007-0430





Report Number:	23-003614/D003.R000	
Report Date:	04/10/2023	
ORELAP#:	OR100028	
Purchase Order:		
Received:	04/04/23 13:38	



Page 8 of 9
Test results relate only to the parameters tested and to the samples as received by the laboratory. Test results meet all requirements of NELAP and the Columbia Laboratories quality assurance plan
unless otherwise noted. This report shall not be reproduced, except in full, without the written consent of this laboratory. Samples will be retained for a maximum of 30 days from the receipt date unless
prior arrangements have been made.
Testing in accordance with: OAR 333-007-0430





23-003614/D003.R000 **Report Number: Report Date:** 04/10/2023 **ORELAP#:** OR100028 **Purchase Order: Received:** 04/04/23 13:38

Explanation of QC Flag Comments:

Code	Explanation
Q	Matrix interferences affecting spike or surrogate recoveries.
Q1	Quality control result biased high. Only non-detect samples reported.
Q2	Quality control outside QC limits. Data considered estimate.
Q3	Sample concentration greater than four times the amount spiked.
Q4	Non-homogenous sample matrix, affecting RPD result and/or % recoveries.
Q5	Spike results above calibration curve.
Q6	Quality control outside QC limits. Data acceptable based on remaining QC.
R	Relative percent difference (RPD) outside control limit.
R1	RPD non-calculable, as sample or duplicate results are less than five times the LOQ.
R2	Sample replicates RPD non-calculable, as only one replicate is within the analytical range.
LOQ1	Quantitation level raised due to low sample volume and/or dilution.
LOQ2	Quantitaion level raised due to matrix interference.
В	Analyte detected in method blank, but not in associated samples.
B1	The sample concentration is greater than 5 times the blank concentration.
B2	The sample concentration is less than 5 times the blank concentration.

 Www.columbialaboratories.com
 Page 9 of 9

 Test results relate only to the parameters tested and to the samples as received by the laboratory. Test results meet all requirements of NELAP and the Columbia Laboratories quality assurance plan unless otherwise noted. This report shall not be reproduced, except in full, without the written consent of this laboratory. Samples will be retained for a maximum of 30 days from the receipt date unless prior arrangements have been made.

 Testing in accordance with:
 OAR 333-007-0430